

SP-VIS 200

SP-VIS 200 Series UV-Vis Spectrophotometer



Operating Manual

SPECTRUM INSTRUMENTS AUST PTY LTD.

**Add: 1211/ 7, Riverside Quay, South Bank,
Victoria 3006, Australia.**

TEL: +61 3 90415554, FAX: +61 3 90415554

Email: info@spectrum-instruments.com.au

**General information about Spectrum Instruments Aust Pty Ltd.
on the internet; <http://www.spectrum-instruments.com.au>**

Copyrights and Trademarks

Microsoft, Windows 2000, Windows XP, Windows NT, MS Excel
are registered trademarks of Microsoft Corp.

Edition – September 2013

Technical documentation made by: Spectrum Instruments

This documentation describes the state of this product at the time of publishing. It needs not necessarily agree with future versions of the product. Subject to change!

© Copyright 2013 Spectrum Instruments Aust Pty Ltd.

Content	Page
1 Introduction.....	4
1.1 Warning and safety labels on SP-VIS 200.....	4
1.2 Notes on the User's Manual.....	5
2 Safety notes.....	6
3 Disposal	8
4 Performance Specifications.....	8
5 Design of the SP-VIS 200.....	9
6 Unpacking instructions.....	10
6.1 Installation and transport conditions.....	10
6.2 Installation and start-up.....	10
6.3 Select light source.....	11
6.4 Turn the Lamps on/off.....	11
7 Maintenance.....	12
7.1 Replacing Halogen lamp.....	12
7.2 Replacing Deuterium lamp.....	13
8 Accessories.....	15
9 Display and Buttons of SP-VIS 200.....	17
10 Software function.....	18
11 Connect to PC.....	24

1 Introduction

The SP-VIS 200 is a single beam spectrophotometer. The instrument is compact, easy-to-use, and ruggedness. The large sample compartment can accept various accessories and cell from 10mm to 100mm path length. The built-in RS-232 serial interface enables the instrument to be connected to a printer.

The SP-VIS 200 is ideal to use in a clinical laboratory, biochemical laboratory, petrochemical laboratory, environmental protection and various fields of quality control. This user's manual provides the description of the SP-VIS 200, a list of main specifications and available accessories.

After carefully unpacking the content, check the materials against the packing list to ensure that you have received everything in good condition. If some parts are missing, damaged or have any other defects, please contact you dealer or the sales representative immediately.

1.1 Warning and safety labels on SP-VIS 200

The following warning and safety labels are located on the SP-VIS 200:



Caution! Hazardous area. Do not remove this cover under any circumstances. Trained service peoples only. No serviceable components inside. Select the right voltage.

Disconnect the power cable before opening the device cover or replacing fuses!

1.2 Notes on the User's Manual

In this manual, the following symbols are used to refer you to warnings and notes:



Warning!

Warning messages alert you to a specific procedure or practice which, if not followed correctly, could cause personal injury.



Caution!

Caution messages refer you to procedures which, if not observed, could result in damage to the equipment.



Caution! Dangerous electric voltage!



Caution! Emission of UV radiation!



Disconnect power cable!



Note

This note must be followed to obtain correct measurement results.

For easier navigation within the manual, the manual uses the following system:

- Chapters and illustrations are numbered consecutively.
- Every illustration has its own caption.
- Steps of operation are numbered.
- Cross-references are marked by an arrow (e.g. →Notes on the User's Manual p.5)

2 Safety notes

For your personal safety and for a trouble-free operation, carefully read this chapter before starting up the SP-VIS 200.

Observe all safety notes given in this manual and all messages and notes displayed by the control software on the screen.

In addition, observe the safety notes for system components of other manufacturers (e.g. PC, printer) supplied together with the device.

In particular, you should also observe the safety notes given on the labels and the information on the use, storage and disposal of reagents, reaction cells and the notes on the package inserts of test kits.



Intended use!

The SP-VIS 200, including original accessories, may only be used for the applications described in this manual. The manufacturer cannot assume any liability for any other application, including that of individual modules or single parts. This also applies to all service or repair work that is not carried out by authorized service personnel. All warranty claims shall be forfeited.



Local safety regulations!

Observe the local safety regulations relevant to the operation of the device (e.g. labor safety regulations, regulations for the prevention of and protection against accidents).

References to possible risks given in this manual do not replace the relevant labor safety regulations that are to be followed.



Personnel!

The SP-VIS 200 may only be operated by appropriately qualified and trained personnel.

For that, the knowledge of the information provided by this manual is indispensable and taken for granted.



Emergency stop!

In the case of emergency, disconnect the SP-VIS 200 and its components from line power by removing the power plug from the power outlet.

Caution! This procedure involves the risk of data loss and damage to the operating system!



Electric shocks!

The SP-VIS 200 is electrically powered. At several parts inside the system, **extremely hazardous electrical** voltages are accessible!

To ensure Protection Class I (protective earth connection) of the device, connect the power plug of the device only to a power outlet with protective earth conductor. The protective effect must not be undone by the use of an extension cable without protective earth conductor.

Before opening the device or removing any device covers, disconnect the power cable!

Before connecting the SP-VIS 200 to the power outlet, make sure the line voltage supplied agrees with the operating voltage specified on the rating plate at the rear panel of the device. Operation of the device with any other operating voltage than that specified may result in the destruction of the device.

Only use fuses of the specified type.



Do not operate the device in explosion-risk rooms!



UV radiation!

Never look directly or indirectly, e.g. via a mirror, into the radiation emitted by the lamp to avoid the risk of getting conjunctivitis!

Therefore, switch off the lamp for lamp replacement!



Lamp damage!

Do not grasp functioning halogen and deuterium lamps on their glass bulbs.

Contamination on the bulb affects the radiation properties of the lamps.

Especially protect the beam-exit window of the deuterium lamp.



Accumulation of heat

Heat accumulation might result in overheating and faults on the device.

Make sure not to cover the ventilation slots of the SP-VIS 200!

Before you switch on the SP-VIS 200, remove its dust cover!



Handling liquids

Take care to ensure that no liquids can enter the SP-VIS 200 to avoid any damage to the device.



Risk of corrosion

Do not set up the device in the vicinity of aggressive vapors, e.g. strongly corrosive acid or caustic vapors!

The vapors might corrode the connections, mechanical and optical components of the device.

Avoid exposing the sample compartment to strongly corrosive substances! For the analysis of such samples, use vapor-tight stoppered cells.

After the measurement, do not leave the samples unnecessarily long in the sample compartment of the device!

3 Disposal

The owner/operator of the SP-VIS 200 must ensure to dispose of the waste material (samples) cropping up in the analyses in accordance with the relevant legal and local regulations.

When the service life of the device is over, dispose of the device with its electronic components as electronic scrap observing the relevant regulations.

4 Performance Specifications

Model	SP-UV 200	SP-VIS 200
Optical system	1200 grooves/mm diffraction grating monochromator.	1200 grooves/mm diffraction grating monochromator.
Optic type	Single beam	Single beam
Light path	10mm-100mm	10mm-100mm
Spectral bandwidth	2nm	4nm
Wavelength range	190-1100nm (0.1nm increment)	320-1100nm (0.1nm increment)
Wavelength accuracy	$\pm 0.5\text{nm}$	$\pm 0.5\text{nm}$
Wavelength repeatability	$\pm 0.3\text{nm}$	$\pm 0.3\text{nm}$
Stray light	$\leq 0.05\%T$ (at 220nm NaI & 340nm, NaNO ₂)	$\leq 0.3\%T$ (at 340nm, NaNO ₂)
Photometric mode	T, A and Conc. measurement	T, A and Conc. measurement
Photometric range	0-125.0%T, -1.000-4.000A*, 0.000-9,999 Conc.	0-125.0%T, -1.000-4.000A*, 0.000-9,999 Conc.
Photometric accuracy	$\pm 0.3\%T$	$\pm 0.3\%T$
Photometric repeatability	$\pm 0.2\%T$	$\pm 0.2\%T$
Baseline stability	$\pm 0.002A/hr$ (@500nm, after lamp turn on for 1 hour)	$\pm 0.002A/hr$ (@500nm, after lamp turn on for 1 hour)
Baseline flatness	$\pm 0.002A$	$\pm 0.003A$
Noise level	$\leq 0.001A$ (at 0A, 500nm)	$\leq 0.001A$ (at 0A, 500nm)
Light source	Tungsten-Halogen & Deuterium	Tungsten-Halogen
Detector	Silicon Photodiode	Silicon Photodiode

*When the instrument is connected to the PC and controlled by software.

5 Design of the SP-VIS 200

Front side



Fig.1 Front side of SP-VIS 200

Back side



Fig. 2 Back side of SP-VIS 200

6 Unpacking instructions

6.1 Installation and transport conditions

Installation conditions

Considering the dimensions of the SP-VIS 200 and a clearance of approximately 100mm to the sides and the front, the device needs a bench space of at least 665 x 565 mm.



Caution!

Do not cover the ventilation slots of the device!

Do not place any objects on the device!

Do not operate the device in explosion-risk rooms!

Keep aggressive vapors, e.g. strongly corrosive acid or caustic vapors away from the vicinity of the device.

The installation site of the SP-VIS 200 must meet the following requirements:

- The place should be free of draft, dust, corrosive vapors as well as vibrations.
- Do not set up the SP-VIS 200 near electromagnetic fields (e.g. of motors).
- Avoid dripping or splashing water near the instrument.
- Do not expose the device to direct sunlight or the radiation of heaters.
- Temperature range: + 15 ... + 35 °C
- Humidity: up to 85 % (at + 30 °C)

Conditions for storage and transport

For transport and storage, the following ambient conditions must be complied with:

- Temperature range in transport: -40...+70 °C
- Relative humidity in transport: up to 95 %

6.2 Installation and start-up



Observe correct position of device!

Move the SP-VIS 200 only in upright position (observe label on transport case).

Do not turn the device upside down.



Electric shock

Before connecting the SP-VIS 200 and the PC to line power, make sure the available line voltage agrees with the operating voltage specified on the type labels. Operation of the devices on line voltages other than the specified operating voltage may result in their destruction.

1. Unpack the contents carefully and check the materials.
 2. Place the instruments in suitable location.
 3. Remove any obstructions or materials that could hinder the flow of air under and around the instrument vibration.
 4. Select the right voltage.
 5. Plug the power cord into an outlet. Turn on the instrument and allow it to warm up for 20 minutes before taking any readings. A self-test starts automatically.
- Your spectrophotometer is now ready for operation.

6.3 Select light source

You can select light source switch point between 325nm to 355nm.

1. Press [Function]. (Press again, if necessary, to display ***lamp Exchange*** menu item).
2. Press [4], to select ***lamp Exchange*** menu item. Press [Enter].
3. Use [0] ~ [9] to set light source switch point. Press [Enter].

6.4 Turn the Lamps on/off

You can turn off halogen lamp or the deuterium lamp, when your methods do not need it.

1. Press [Function]. (Press again, if necessary, to display ***Lamp on/off*** menu item).
2. Press [5], to select ***Lamp on/off*** menu item. Press [Enter].
3. Press [1], to turn on tungsten lamp. Press [2], to turn off tungsten lamp. Press [3], to turn on deuterium lamp. Press [4], to turn off deuterium.

7 Maintenance

This SP-VIS 200 spectrophotometer has been designed to be durable and reliable.

Therefore, the routine maintenance of the instrument is minimal.

However, you will need to replace the lamp or you may want check the wavelength calibration and photometric linearity to verify the instrument's performance.



The manufacturer recommends only authorized service representatives to perform procedures requiring removal of the instrument's cover and replacement of electrical components. To protect both of yourself and your instrument, be sure to contact the authorized service representative to perform any service procedure, when you do not feel comfortable performing.

7.1 Replacing Halogen lamp



Electric shock!

Before changing the halogen lamp, make sure to switch the device off with the power switch.



Caution: Risk of burns!

Allow the halogen lamp to cool down to avoid the risk of burns!



Caution: Eye injury!

Never look into the lighting lamp as the radiation may cause eye injury!



Caution!

Do not touch the glass bulb of the lamp to avoid any contamination affecting its performance!

1. Turn the power off and unplug the instrument.
2. Put all samples and not fixed elements out of the sample compartment.
3. Place the SP-VIS 200 carefully on a soft underlay on the top side of the device so that you can access the underside.




Fig. 3 Lamp house cover on the underside of SP-VIS 200

- Remove the two screws (11) and the lamp house cover (12) on the underside of the instrument.




Fig. 4 Lamp compartment of SP-VIS 200, opened

- Unscrew the holder of the halogen lamp (13) and take the holder with the halogen lamp out of the device.




Fig. 5 SP-VIS 200 holder of the halogen lamp

- Remove the lamp (14) from lamp holder
- Replace the new lamp carefully in the holder.



Caution

Do not handle Halogen lamp with bare fingers. Use a piece of tissue or cloth when handling the lamp. The oil from your fingers can cause the lamp to burn out prematurely!

- Replace and fix the holder with the new halogen lamp in the device
- Close the cover of the lamp house using the 2 screws.
- Replace the device with the top side on top.

7.2 Replacing Deuterium lamp



Risk of electric shock!

Before changing the deuterium lamp, make sure to switch the device off with the power switch.



Caution: Risk of burns!

Allow the deuterium lamp to cool down to avoid the risk of burns!



Emitted UV radiation!

Never look directly or indirectly, e.g. via a mirror, into the radiation emitted by the lamp to avoid the risk of getting conjunctivitis!



Dirt on beam exit window!

Do not touch the glass bulb of the new deuterium lamp. Particularly avoid contamination on the quartz glass beam exit window! If you touched the glass bulb with your fingers, wipe the fingerprints off with a clean, lint free cloth and pure alcohol.

The performance of the lamp will deteriorate by contamination.

1. Turn off the power and unplug the instrument
 2. Put all samples and not fixed elements out of the sample compartment
 3. Place the SP-VIS 200 carefully on a soft underlay on the top side of the device so that you can access the underside
 4. Remove the two screws (11) and the lamp house cover (12) on the underside of the instrument
 5. Unscrew the 2 screws of the deuterium lamp (13), disconnect the electronic plug connection (15) and take the deuterium lamp out of the device
-



Do not handle Deuterium lamp with bare fingers. Use a piece of tissue or cloth when handling the lamp. The oil from your fingers can cause the lamp to burn out prematurely!

6. Replace the new deuterium lamp and fix it with the screw
7. Close the cover of the lamp house using the 2 screws.
8. Replace the device with the top side on top.

8 Accessories

For the installation of accessory use the two screws on the bottom of the sample compartment.

8.1 Manual 4 cell changer

The standard configuration is a 4 cell changer for 1 cm cells. There is also 4 cell changer for 5 and 10 cm cell available. As an option there are also 4 cell changer for cells with 50 and 100 mm path length available.




Fig. 6 Manual 4 cell changer

Use the cell selector arm for the manual cell changer to set the position of the manual cell changer.




Fig. 7 Positions of the 4 cell changer

8.2 Adjustable cell holder

Micro cells and ultra-micro cells have the standard outside width of 12.5 mm, but a reduced inside width or filling aperture of 1 to 4 mm. This requires accurate alignment of the cell in the light beam.

This adjustable cell holder can be used for cells with an aperture high of 8.5 mm.




Fig. 8 SP-VIS 200 adjustable cell holder

Insert the adjustable cell holder in the sample compartment with screw 1 to the back side. Use the screw 1 to adjust the cell horizontal to the light beam and screw 2 to adjust the cell vertical to the light beam. The cell has the optimal position if the highest transmittance is reached.

8.3 Round cell holder

This special cell holder can be used for round cell of standard test kits. Cells with a diameter up to 16 mm and a high up to 11 cm can be used.



Fig. 9 SP-VIS 200 Round cell holder

8.4 Automatic cell changer

The automatic cell changer accommodates 6 cells a 10mm path length cell of 12.5 x 12.5 x 45 (L x W x H in mm).




Fig.10 Automatic 6 cell changer

To install this accessory in the sample compartment of the SP-VIS 200 switch the device off, plug the electrical connection in the internal serial port at the back site of the sample compartment. Use two screws to install the system. Start the instrument new.

8.5 Printer

To install the printer to the SP-VIS 200 switch the device off, plug the electrical connections into the USB at the back side of the device. Start the instrument new.

9 Display and Buttons of SP-VIS 200




Fig. 11 Display and buttons

Display

The display shows details of what is currently happening in the spectrometer and allows you to make changes to make changes to parameters as necessary.

Keypad

All of the parameters required to operate the SP-VIS 200 can be entered through the spectrometer's keypad. Normally, however, most of these parameters are controlled using an attached data system, such as Spectrum's Win-SP software.

When a Spectrum data system is controlling your SP-VIS 200, it is possible for the data system to disable editing of the spectrometer's current method from the keypad.

10 Software function

10.1 Function

Press [Function] to call menu.

- | | |
|--------------------------|--|
| 1. Abs/%T/Conc. | Absorbance/Transmittance/Concentration measurement |
| 2. 0%T Dark | Currency Correction |
| 3. Factor/Conc. | Concentration factor edit |
| 4. Lamp Exchange | Set light source switch point |
| 5. Lamp on/off | Turn off light source |
| 6. Standard Curve | Create a standard curve |
| 7. Language | Select display language. |
| 8. Com. with PC | Connect with PC |
| 9. Set time | Set the date and time |

10.2 Select Language Mode

1. Press [Function]. (Press again, if necessary, to display **Language** menu item). The LCD will display as blow.

- 5. Lamp on/off
 - 6. Standard Curve
 - 7. Language
 - 8. Com. With PC

2. Press [7], to select **Language** menu item. Press [Enter].The LCD will display as blow.

- Select Language:

 - 1. Chinese
 - 2. English
 - 3. German

3. For Chinese press [1], for English press [2], for German press [3]

10.3 Turn the Lamps on/off

You can turn off halogen lamp or the deuterium lamp, when your methods do not need it.

1. Press [Function]. (Press again, if necessary, to display **Lamp on/off** menu item).The LCD will display as blow.

- 5. Lamp on/off
 - 6. Standard Curve
 - 7. Language
 - 8. Com. With PC

2. Press [5], to select **Lamp on/off** menu item. Press [Enter].The LCD will display as blow.

1. VIS lamp on
 2. VIS lamp off

3. Press [1], to turn on tungsten lamp. Press [2], to turn off tungsten lamp.

10.4 Absorbance / Transmittance / Concentration Measurement

1. Press [GoTo WL], select wavelength.

2. Press [Function]. (Press again, if necessary, to display **Abs/%T/Conc.** menu item).The LCD will display as blow.

1. Abs/%T/Conc
 2. 0%T
 3. Factor/Conc
 4. Lamp Exchange

3. Press [1], to select Abs/%T/Conc. menu item. Press [Enter].The LCD will display as blow.

1. Abs
 2. Tran
 3. Conc

4. Measurement : Absorbance/Transmittance Measurement

1) Press [1], to select Abs menu item. The LCD will display as blow.

546.0 nm ABS

- 2) Place the cell with blank solution in the blank position (position 1) of the cell holder.
- 3) Place the samples or standard solution in the other positions of the cell holder.
- 4) Close the sample compartment cover. Pull the cell holder's rod with the blank sample in the measurement beam. Measure the blank value by pressing the [Auto Zero] until display shows "100.0%T" or 0.000Abs.
- 5) Pull the cell holder's rod to the next sample position and read the %T or Abs value of the samples.

Note

Pay attention that second step of the cell changer is only for reading 0%T. There are two steps between Position 1 and 2 of the manual cell changer.

Alternatively, if you use a single cell holder, at first put the blank solution in the light beam and set the value to 100%T or 0Abs by pressing the [Auto Zero]. Then put the sample or standard cell in the cell holder

6) When the printer is connected, you can print out the result by pressing [P/C].

Note

Choose cells with the same path length for all blanks, standards and samples of the analytical method you are using.

For measurements below 325 nm, use quartz cells. Glass or plastic cells are only usable above 325 nm.

The liquid level must be at least 20 mm in a standard square cell.

10.5 Concentration mode

10.5.1 Determination of the concentration via one point

calibration

The results are calculated with the following equations:

$$\text{CONC} = \text{ABS} * \text{CONC}_{\text{Standard}} / \text{ABS}_{\text{Standard}}$$

1. Press [GoTo WL], select wavelength.
2. Place the blank solution in the sample compartment and close the sample compartment cover.
3. Press [Auto Zero] to measure the blank solution.
4. Remove the blank solution from the sample compartment.
5. Press [Function]. (Press again, if necessary, to display **Factor/Conc** menu item). The LCD will display as blow.

1. Abs/%T/Conc
2. 0%T
3. Factor/Conc
4. Lamp Exchange

6. Press [3] to select Factor/Conc. menu item. Press [Enter].The LCD will display as blow.

1. Std. FACT
2. Std. Conc

7. Press [2] to select **Std. Conc.** menu item. Press [Enter]. The LCD will display as blow.

Input
Std. Conc 1.00

8. Insert the standard solution into the cell holder or cell changer. Close the sample compartment cover.
9. Use digital key to enter the concentration of the standard, press [Enter].
10. Insert the sample solution into the cell holder or cell changer. Close the sample compartment cover.
11. Read the results directly in concentration units on the digital display.

10.5.2 Determination of the concentration via factor

The results are calculated with the following equations:

$$\text{CONC} = \text{ABS} * \text{FACTOR}$$

1. Press [GoTo WL], select wavelength.
2. Place the blank solution in the sample compartment and close the sample compartment cover.
3. Press [Auto Zero] to measure the blank solution.
4. Remove the blank solution from the sample compartment.
5. Press [Function]. (Press again, if necessary, to display **Fact/Conc** menu item). The LCD will display as blow.

1. Abs/%T/Conc
 2. 0%T
 3. Factor/Conc
 4. Lamp Exchange

6. Press [3] to select Factor/Conc. menu item. Press [Enter]. The LCD will display as blow.

1. Std. FACT
 2. Std. Conc

7. Press [1] to select **Std. FACT** menu item. Press [Enter]. The LCD will display as blow.

Input
Std. FACT 1.00

8. Use digital key to enter the factor of the standard, press [Enter].
9. Insert the sample solution into the cell holder or cell changer. Close the sample compartment cover.
10. Read the results directly in concentration units on the digital display.

10.5.3 Determination of the concentration via calibration curve

The results are calculated with the following equations:

$$\text{CONC} = \text{ABS} * B + K$$

1. Place the blank solution in the sample compartment and close the sample compartment cover.
2. Press the **[Auto Zero]** button to zero the blank solution.
3. Remove the blank solution from the sample compartment.
4. Press [Function]. (Press again, if necessary, to display **Standard Curve** menu item).The LCD will display as blow.

5. Lamp on/off
 6. Standard Curve
 7. Language
 8. Com. With PC

5. Press [6] to select **Standard Curve** menu item. Press [Enter].The LCD will display as blow.

- 1. Create Curve
- 2. Load Curve
- 3. Delete Curve
- 4. Print Curve

6. Press [1] to select **Create Curve** menu item. Press [Enter].The LCD will display as blow.

- 1. No. of Samples
- 2. Unit
- 3. Input Conc
- 4. Test Sample

7. Press [1] to select **No. of samples** menu item. Press [Enter].The LCD will display as blow.

Input
Sample Num:

8. Use digital key to enter samples number, press [Enter].

9. Press [2] to select **Unit** menu item. Press [Enter].The LCD will display as blow.

- 1. mg/L
- 2. ug/L
- 3. ng/kg
- 4. ug/kg

10. Use digital key to enter sample unit, press [Enter].

11. Press [3] to select **Input Conc** menu item. Press [Enter].The LCD will display as blow.

Input
Conc:
Number:1

12. Use digital key to enter sample concentration, press [Enter].

13. Press [4] to select **Test Sample** menu item. Press [Enter].The LCD will display as blow.

Number: 1
Conc: 1.00
Abs.:

14. Insert the standard solutions in the light beam. Press [Enter].

If all standards are measured, you get the coefficient of the calibration K and B as well as the coefficient of determination R.

10.5.4 Use an existing calibration curve

1. Press [Function]. (Press again, if necessary, to display **Standard Curve** menu item).The LCD will display as blow.

- 5. Lamp on/off
- 6. Standard Curve
- 7. Language
- 8. Com. With PC

2. Press [6] to select **Standard Curve** menu item. Press [Enter]. The LCD will display as blow.

- 1. Create Curve
- 2. Load Curve
- 3. Delete Curve
- 4. Print Curve

3. Press [2] to select **Load Curve** menu item. Press [Enter]. The LCD will display as blow.

Curve No.:

4. Use digital key to load curve stored in spectrometer, press [Enter].

10.5.5 Delete an existing calibration curve

1. Press [Function]. (Press again, if necessary, to display **Standard Curve** menu item).The LCD will display as blow.

- 5. Lamp on/off
- 6. Standard Curve
- 7. Language
- 8. Com. With PC

2. Press [6] to select **Standard Curve** menu item. Press [Enter]. The LCD will display as blow.

- 1. Create Curve
- 2. Load Curve
- 3. Delete Curve
- 4. Print Curve


3. Press [3] to select **Delete Curve** menu item. Press [Enter]. The LCD will display as blow.

Curve No.:


4. Use digital key to delete curve stored in spectrometer, press [Enter].


10.6 Set time

1. Press [Function]. (Press again, if necessary, to display **Set time** menu item). The LCD will display as below.



2. Press [9], to select **Set time** menu item. Press [Enter].The LCD will display as blow.




3. For example: The data is 2014/01/10, time is 16:20:10, you can input 14.01.10. After that the symbol () move to [Time:], then input 16.20.10 and press [Enter]. The LCD will display the date and time about 10 seconds and change automatically to the main page.


11. Connect to PC

You can connect to a personnel computer (PC) when you need to control the system by PC control.

1. Press [Function]. (Press again, if necessary, to display **Com. With PC** menu item).The LCD will display as blow.



2. Press [8], to select **Com. With PC** menu item. Press [Enter].The LCD will display as blow.



3. Press [1], to use RS232 connect with PC. Press [2], to use USB connect with PC.



SPECTRUM INSTRUMENTS AUST PTY LTD.

**Add: 1211/ 7, Riverside Quay, South Bank,
Victoria 3006, Australia.**

TEL: +61 3 90415554, FAX: +61 3 90415554

Email: info@spectrum-instruments.com.au

<http://www.spectrum-instruments.com.au>